

HCV Rapid Kit Storage and Testing Room Temperature Log

Thermometer location: _____

Acceptable temperature range*: 15°C to 30°C or 59°F to 86°F

Month/Year: _____

Day	Temperature	Initials	Day	Temperature	Initials
1			17		
2			18		
3			19		
4			20		
5			21		
6			22		
7			23		
8			24		
9			25		
10			26		
11			27		
12			28		
13			29		
14			30		
15			31		
16					

***HCV rapid test kit acceptable range:**

- Test kit storage is 2° to 30°C or 36° to 86°F (taken daily)
- Testing area is 15° to 37°C or 59° to 99°F (taken on testing days)

NOTE: Periodically (e.g., every six months) check thermometer performance and document.

Corrective Action

Date	Action Taken	Initials

Reviewed by and date:
